

27 May 1966
552 - OD-292
LHB:rf

STAT

Subject: [redacted] Progress Report - April
1966 Projects No. 552 and 552A

Gentlemen:

Enclosed are three (3) copies of [redacted]
Progress Report on Projects 552 and 552 A for the period
April 1966.

STAT

Very truly yours,

[redacted signature box]

STAT

Executive Vice President

Encl: (3) P.R.
Cert. #632636

Declass Review by
NIMA/DOD

552 - OD-290

PROGRESS REPORT
For
VERSATILE, HIGH PRECISION STEREO
POINT TRANSFER DEVICE

Progress Report 552 & 552A for April 1966

Stereo Viewer 552A #103 is now entering the debugging stage. Many of the electrical, mechanical and optical adjustments have been made, but complete system has not been operating, although some debugging has been done. Counters have been received from customer and tested for operation. Measurement runs have not been made.

Acceptance period will be delayed until early June because of prolonged installation work on the 552 Point Transfer Device.

PROGRESS REPORT
For
VERSATILE, HIGH PRECISION STEREO
POINT TRANSFER DEVICE

Period Covered: April 1966
Dated: 27 May 1966
Job No.: #552, #552A
Document No.: OD-290

552 - OD-290

PROGRESS REPORT
For
VERSATILE, HIGH PRECISION STEREO
POINT TRANSFER DEVICE

Progress Report 552 & 552A for April 1966

Stereo Viewer 552A #103 is now entering the debugging stage. Many of the electrical, mechanical and optical adjustments have been made, but complete system has not been operating, although some debugging has been done. Counters have been received from customer and tested for operation. Measurement runs have not been made.

Acceptance period will be delayed until early June because of prolonged installation work on the 552 Point Transfer Device.

PROGRESS REPORT
For
VERSATILE, HIGH PRECISION STEREO
POINT TRANSFER DEVICE

Period Covered: May 1966
Dated: 27 June 1966
Job No.: #552, #552A
Document No.: OD-293

PROGRESS REPORT
For
VERSATILE, HIGH PRECISION STEREO
POINT TRANSFER DEVICE

Progress Report 552 & 552A for May 1966

Stereo viewer 552A #103 is now into the final adjustment and checkout stages. Scanning drives are completely aligned with optics alignments to be finished early in June. Encoders and measurement capability have been checked with measurement repeat precision of 1 count, or 2 1/2 microns. Acceptance for shipment should be expected early in June.

Electrical debugging and adjustment are being completed on 552A #104. Design of modifications has been started although actual mechanism changes should be commenced late in June.

Financial Report Enclosed