

99-261-5

CONFIDENTIAL

Declass Review by NGA.

NPIC/P&DS/D/6-888
11 April 1966

MEMORANDUM FOR: Assistant for Photographic Analysis, NPIC

SUBJECT: Poor Illumination from Modified Tri-Simplex Projectors

REFERENCE: Memorandum IPO/O&B/M-117-66 of 14 March 1966

1. In accordance with the referenced request, the Plans and Development Staff has determined the cause of restricted projection illumination on the modified Tri-Simplex Projectors. By adjustment of the distance between the projector stage platen and the condenser lens, the entire area of interest within the U-shaped base can be illuminated satisfactorily.

2. Coincidentally, as discussed by [redacted] P&DS, and [redacted] PAG, the projector provided by PAG for P&DS investigation contained a broken condenser lens. With this situation corrected, the projector should function properly.

3. The Plans and Development Staff hopes that the service provided is satisfactory and that the Photographic Analysis Group will not hesitate to call upon it again.

[redacted]

Colonel, USAF
Assistant for Plans and Development, NPIC

Distribution:

Orig & 1 - Addressee

- 1 - JAP
- 2 - DB/P&DS

NPIC/P&DS/DB: [redacted] (11 Apr 66)

CONFIDENTIAL

GROUP 1
Excluded from automatic
downgrading and
declassification