

CONFIDENTIAL

Approved For Release 2005/05/02 : CIA-RDP78B04770A002300010010-6

Declass Review by NGA.

NPIC/P&DS/D/6-1487
28 July 1966

MEMORANDUM FOR THE RECORD

SUBJECT: Twin Light Source Light Table with Measuring Stage
Project #997546 (02107)

1. On this date the undersigned discussed the following items with [redacted] IAD/CIA and agreement was reached to incorporate the listed features into the subject instrument.

25X1

A. The stage counters should read either in feet *.0001"/in* or metric units. *5 AUG*

B. A specification as to "the accuracy over what distance traveled" should be stated.

C. The stage travel in both axes should be specified. *3"x3"*

D. The microscope support slide should move four inches *✓* vice the three inches proposed.

E. The "split nut" coupling device for engaging the measuring stage should be described. *This item will be changed to a screw with 10 threads/in*

5 AUG 66

2. On this same date, the undersigned discussed the above points with [redacted] during his visit to NPIC. He agreed to investigate and send his written reply as soon as possible.

25X1

[redacted]

25X1

Exploitation Systems Section, P&DS

Distribution:

- Original - Project #997546
- 2 - Chrono/DB

clc

Approved For Release 2005/05/02 : CIA-RDP78B04770A002300010010-6

CONFIDENTIAL

