

SECRET

02107

CONFIDENTIAL

UNCLASSIFIED

Approved For Release 2005/05/02 : CIA-RDP78B04770A002300010015-1
CONTRACT INSPECTION REPORT

15-6K NO. 25X

TO:
CONTRACT ADMINISTRATION & SETTLEMENT
BRANCH/PD/OL

DATE
4 November 1966
INSPECTION REPORT NO. (If final, so state)
1
ESTIMATED COMPLETION DATE
16 January 1967

NAME OF CONTRACTOR

Declass Review by NGA.

TYPE OF COMMODITY OR SERVICE
Twin Light Source Table With Measuring Stage

THE CONTRACTOR IS ON SCHEDULE
 YES NO
PER CENT OF WORK COMPLETED - 10%
PER CENT OF FUNDS EXPENDED - Fixed Price

THE CONTRACTOR WILL PROBABLY REMAIN WITHIN ALLOCATED FUNDS
 YES NO IF ANSWER IS "NO" ADVISE RECOMMENDATION AND/OR ACTION OF SPONSORING OFFICE, ON REVERSE HEREOF. IF KNOWN, INDICATE MAGNITUDE OF ADDITIONAL FUNDS INVOLVED.

HAS AN INTERIM REPORT, FINAL REPORT, PROTOTYPE, OR OTHER END ITEM BEEN RECEIVED FROM THE CONTRACTOR DURING THE PERIOD? YES NO (If yes, give details on reverse side.)
HAS GOVERNMENT-OWNED PROPERTY BEEN DELIVERED TO CONTRACTOR DURING THIS PERIOD? YES NO (If yes, indicate items, quantity, and cost on reverse side.)

INCENTIVES

IS THIS AN INCENTIVE CONTRACT IF YES, CHECK TYPE
 COST AWARD FEE PERFORMANCE DELIVERY
 YES NO

NOTE:
USE REVERSE SIDE FOR COMMENTS.
FINAL REPORT MUST CONTAIN INCENTIVE EVALUATION.

OVERALL PERFORMANCE OF CONTRACTOR

1. OUTSTANDING 4. ABOVE AVERAGE 7. UNSATISFACTORY
2. EXCELLENT 5. AVERAGE
3. VERY GOOD 6. MINIMUM ACCEPTABLE
IF OVERALL PERFORMANCE OF CONTRACTOR IS UNSATISFACTORY OR MINIMUM ACCEPTABLE INDICATE REASONS ON REVERSE SIDE.

RECOMMENDED ACTION

CONTINUE AS PROGRAMMED WITHHOLD PAYMENT PENDING SATISFACTORY PERFORMANCE
 TERMINATE OTHER (Specify)
IF TERMINATION IS RECOMMENDED OR IF THIS IS A FINAL REPORT PUT COMMENTS ON REVERSE IN NARRATIVE FORM ON CONTRACTOR'S PERFORMANCE AND CERTIFY THAT ALL DELIVERABLE ITEMS UNDER THE CONTRACT HAVE BEEN RECEIVED. THESE INCLUDE, WHERE APPLICABLE, THE FOLLOWING:

ITEM	REC'D	DOES NOT APPLY	ITEM	REC'D	DOES NOT APPLY
PROTOTYPES			MANUALS		
DRAWINGS AND SPECIFICATIONS			FINAL REPORT		
PRODUCTION AND/OR OTHER END ITEMS			SPECIAL TOOLING		
			OTHER GOVERNMENT PROPERTY		

DATE OF LAST CONTACT WITH CONTRACTOR
26 October 1966

SIGNATURE OF INSPECTOR

DIVISION

INSPECTOR

Approved For Release 2005/05/02 : CIA-RDP78B04770A002300010015-1

Inspection visit of 26 October 1966.

1. attendees:

- 1. [redacted]
- 2. [redacted]

25X1

2. The design of the light table was reviewed and the following items were discussed.

a. Items which [redacted] requested to be changed (for their benefit) and the technical monitor agreed.

25X1

- 1. The brightness range will be 200 to 4000 ft. Lamberts operating at a maximum film plane temperature of 95°F in a 70°F ambient. The lamp life will be about 10,000 hours and at maximum brightness the lamp will require approximately 60 milliamps.
- 2. The positions of the counters have been moved to the extreme outside of the instrument, but remain on the same supporting surface.
- 3. The diameter of the stage transport handwheels have been increased to 3 inches.
- 4. Both the overall length and overall width have increased by 2 inch.

b. Requested changes that were not permitted.

- 1. Moving the "X" counter and handwheel onto the moving stage. This would require the counter and handwheel to move.
- 2. Reduction of the speed of motion of the stage. 30 turns of the handwheel will now be required to move the stage 3 inches in either axis.
- 3. Elimination of the divisions on the last drum counter.
- 4. Substitution of [redacted] support ring for the proposed versatile ring. The versatile ring will be retained.

25X1

c. The following measurements will be provided to [redacted]

- 1. The maximum cover glass thicknesses.
- 2. The working distances of the microscopes.

25X1