

NPIC/D-171-70

10 JUN 1970

MEMORANDUM FOR: Assistant Deputy Director for Intelligence *ELW*

SUBJECT : Request for Approval of a Contract for a  
Color Standard Project with [redacted]  
[redacted] from FY-1970 Funds

25X1

25X1

1. This memorandum requests approval for the commitment of R&D funds for a NPIC contract. The specific request is stated in paragraph 9.

2. The advent of high definition color film as an operational acquisition material has necessitated the development of color image evaluation standards and techniques. Methods and equipment suitable for analysis of black and white imagery require extension and modification before they may be applied to color imagery. For certain more routine functions, e.g., chemical process control and film exposure requirements, standard operating conditions and instrumentation have already been established. This is not the case for applications such as the determination of object color from measurements made on the fine detail of color imagery, the measurement of color modulation transfer function and power spectral density, micro-image resolution, and other color image quality parameters. In addition, the response of color film layers during the scanning operation in digital image manipulation must be quantitatively determined.

3. As a prerequisite to making any of the measurements mentioned above, standard instrumentation and data collection techniques must be developed. For image evaluation and manipulation applications, most of the data required can be collected by an instrument called a microdensitometer. Because the measurements made are concerned with image areas on the order of several microns (millionths of a meter) the microdensitometer itself must be calibrated extremely accurately so that variations in its output data can be assigned the proper cause, i.e., electrical, mechanical, optical or photographic.

**SECRET**

GROUP 1 Excluded from automatic downgrading and declassification
---

SUBJECT: Request for Approval of a Contract for a  
Color Standard Project with [redacted]  
[redacted] From FY-1970 Funds

25X1  
25X1

4. Currently, a microdensitometer for measuring black and white and color film is available in Applied Photo-Science Division (APSD). This instrument was designed in 1965 for use with materials then available. While an excellent instrument for black and white image analysis, precalibration experiments on current color imagery have shown that (1) new optics will be required and (2) electronic, mechanical, and optical calibration standards must be developed. As currently designed the objective lenses on the microdensitometer do not have adequate depth of field to focus the three images in a single image position as required. As a consequence, the output data are dependent upon, rather than independent of, the focus position. The situation is analogous to an improperly focused color TV set where color fringing can be observed and the objects are obviously not correctly reproduced. After the microdensitometer has been properly calibrated, quality control procedures are necessary to assure proper day-to-day operation.

5. The proposed contract with [redacted] is designed to overcome the problems discussed above. It will be executed in three overlapping phases during a nine-month period. Phase one will be concerned with acquisition and calibration of new objective lenses. Phase two will involve electronic, mechanical and optical system calibration. The optics in this phase will include filters, mirrors, and condensers in concert with the new objective lenses. Phase three will be concerned with the application of the microdensitometer to color transfer function analyses, the formulation of quality control procedures, and the preparation of a color microdensitometer handbook. This handbook will spell out in detail the procedures for use of the instrument and the techniques for analyzing the data collected. Monthly and final reports, the quality control film standards, and the Handbook constitute the deliverable items.

25X1

6. [redacted] has been selected as the sole source for this project because of their intimate knowledge of the APSD Mann 1032T microdensitometer (they installed it originally), and their extensive experience with micro image quality assessment on both black and white and color films. Accordingly, the risk associated with the calibration portion of the project is very low. The adequacy of the objective lenses, however, cannot be determined in detail before the other variables are pinned down. Consequently, the risk involved here is somewhat greater: contingency plans for the redesign or modification of the objectives have been made.

25X1

SUBJECT: Request for Approval of a Contract for a  
Color Standard Project with [redacted]  
[redacted] from FY-1970 Funds

25X1  
25X1

7. Follow-on action is anticipated in the development of more advanced and sophisticated color imagery analysis techniques. The level of such follow-on action cannot now be estimated and will not necessarily be with the same contractor. This project has been coordinated with APSD.

8. The sterility Code, [redacted] is appropriate; the project officer, [redacted] will assign security classification to the individual reports.

25X1  
25X1

9. It is requested that the negotiation with [redacted] for a contract to conduct the program described at a cost not to exceed [redacted] be approved.

25X1  
25X1  
25X1

[redacted signature box]

25X1

ARTHUR C. LUNDAHL  
Director  
National Photographic Interpretation Center

- Attachments:  
1. Proposal  
2. Form 2420

APPROVED: [redacted] 12 June 70  
Assistant Deputy Director for Intelligence Date

25X1

- Distribution:  
Original - NPIC/TSSG/SC&PS (After Approval)  
1 - ADDI  
2 - NPIC/ODIR  
1 - NPIC/TSSG  
1 - NPIC/TSSG/RED

Declassified in Part - Sanitized Copy Approved for Release 2012/08/30 : CIA-RDP79B00873A000800010042-0

REQUEST TO PROCUREMENT DIVISION FOR SERVICES  
(Other than property or building maintenance and repairs)

OFFICE/DIV/BR TSSG/RED/ATB	REQUEST NO. 5500-6115-70	DATE OF REQUEST 26 May	I CERTIFY THAT FUNDS IN THE ESTIMATED AMOUNT OF \$ [redacted] 25X1 AVAILABLE. CHARGE TO FAN 0155-5350
-------------------------------	-----------------------------	---------------------------	--

PROJECT TITLE Color Standards- (Physical)	25X1
---	------

CONTRACTOR (if known)	[redacted]	25X1
-----------------------	------------	------

TYPE OF SERVICE REQUESTED		APPLICABLE ONLY TO REPAIRS AND MODIFICATIONS	
<input checked="" type="checkbox"/> RESEARCH/DEV	TRANSFER OF FUNDS TO OTHER GOVT. AGENCY (specify if applicable)	TECHNICAL INSPECTION IS REQUIRED BY	
<input type="checkbox"/> GRANT		<input type="checkbox"/> RECEIVING DEPOT T & I	<input type="checkbox"/> TECHNICAL MONITOR
<input type="checkbox"/> REPAIR		<input type="checkbox"/> ITEMS TO BE PICKED UP OR <input type="checkbox"/> SERVICES PERFORMED AT:	
<input type="checkbox"/> MAINTENANCE		ITEMS TO BE REDELIVERED TO:	
<input type="checkbox"/> MODIFICATION			

STERILITY	CONTRACT CLASSIFICATION	WORK CLASSIFICATION	HARDWARE CLASSIFICATION	REPORTS CLASSIFICATION
<input checked="" type="checkbox"/> sc 1	SECRET	UP TO TKH	UNCLASS	UP TO TKH

SHORT SUBSTANTIVE TITLE AND/OR DESCRIPTION OF SERVICE TO BE PERFORMED

Test and assist in selection of color corrected microdensitometer objectives\* that will permit accurate density measurements from color film in one scan of all three layers.

Develop techniques and procedures for determining MTF, effective exposure, power spectrum, etc.

Develop quality control and operating procedures for the APSD tri-chromatic microdensitometer.

Prepare detailed operating and quality control handbook for routine operation of the microdensitometer.

\*For use on the APSD Mann Trichromatic Microdensitometer

(See reverse for specific information required on R&D requests.)

DESIGNATION	APPROVAL SIGNATURE	DATE
C/SSD/TSSG/ND	[redacted]	6-16-70

PROCUREMENT DIVISION USE			
DATE RECEIVED IN PD	RECORDED BY	SECTION ASSIGNED TO	NEGOTIATOR

Declassified in Part - Sanitized Copy Approved for Release 2012/08/30 : CIA-RDP79B00873A000800010042-0

installed the microdensitometer concerned, has modified it previously (to improve sampling) and has developed the color image assessment which will be investigated with the microdensitometer when calibrated. 25X1

The instrument concerned is one of two, both installed  25X1

2. PROPOSAL

Attached

3. DELIVERABLE ITEMS

REPORTS REQUIRED 25 NO. OF COPIES  MONTHLY  INTERIM  QUARTERLY  FINAL

HARDWARE (state type and number) (1) objectives (2) (Zeiss, apochromats) designed for use with APSD microdensitometers, 400 mm working distance, single scan for three accurate traces. (2) Microdensitometer color standard measuring targets (Qty) to be determined

OTHER

4. GFE REQUIRED

80 man hours microdensitometer scanning & testing time is required and has been agreed to by C/APSD.

5. SPECIAL INSTRUCTIONS

ref proposal task I, part III par 2c - delete requirement to use color or dye targets. Metric accuracy not affected here.